

CERTIFICATE OF CALIBRATION



CUSTOMER DETAILS			
NAME ADDRESS	FASTSOURCE LTD CALDER VALE ROAD WAKEFIELD		
POST CODE	WF1 5PH		

CERTIFICATE No.02797

MACHINE DETAILS			
INDICATOR MANUF/TYPE	VISHAY / VT300A		
INDICATOR SERIAL No	T12310501		
SCALE TYPE	WEIGHBRIDGE		
BASE WORK MANUF/No	S S S LTD / -		
SOFTWARE VERSION	-		
CAPACITY	50,000 KG		
INCREMENTS	20 KG		
LOCATION	AS ABOVE		


CALIBRATION DETAILS			
NEXT CERTIFICATE DUE DATE OF CALIBRATION TEST WEIGHT CERT NO ENGINEERS NAME	NOVEMBER 2023 02.12.2022 9647 BEN POWELL		
SIGNATURE	Andrew & Saville		

			LINEARITY AND	HYSTERESIS TEST			
			TARE=		TARE=		
	INDICATION		INDICATION		INDICATION		
LOAD	LOADING UNLOADING		LOADING UNLOADING		LOADING UNLOADING		
0 KG 2,000 KG 10,000 KG 20,000 KG 39,020 KG 44,220 KG	0 KG 2,000 KG 10,000 KG 20,000 KG 39,040 KG 44,260 KG						
	DEFINITIVE TEST						
	NO TARE		TARE=		TARE=		
	INDICATION		INDICATION		INDICATION		
LOAD	LOADING	UNLOADING	LOADING	UNLOADING	LOADING	UNLOADING	

REPEATABILITY TEST						
	AS FOUND			DEFINITIVE		
LOAD	IND	IND	IND	IND	IND	IND

DISCRIMINATION/SENSITIVITY TEST					
LOAD	AD DISPLAY WT ADDED DISP CHANG				

ECCENTRIC LOAD TEST					
	AS FOUND		DEFINITIVE		
POSITION	LOAD	IND	LOAD	IND	
1 2 3 4 5 6 7	44,220 KG 44,220 KG 44,220 KG	44,260 KG 44,260 KG 44,260 KG			



CALIBRATION CARRIED OUT IN ACCORDANCE WITH THE U.K.W.F. CODE OF PRACTICE